

TEST REPORT

LAB NO.: (6621)105-0677

(REVISED)

FORM NO.:

DATE IN: Apr 15, 2021 **DATE OUT:** Apr 21, 2021

NO. OF WORKING DAYS:

REVISED DATE: Apr 27, 2021

PAGE 1 OF 3

This report is amendment of and supersedes the previous (6621)105-0677 dated Apr 21, 2021.

OVERALL RATING	
PASS FAIL DATA	X

Vendor:	WUHAN YOUFU	Agent:	/
	INTERNATIONAL TRADE		
	CO.,LTD		
Fabric Supplier/Mill:	/	Factory/Manufacturer:	/
P.O. No.:	/	Style No.:	/
Sample Description:	AAMI PP+PE ISOLATION	Style Description:	/
	GOWN LEVEL 2		
Color:	BLUE	Country of Origin:	1
Claimed Fabric Weight:	/	Claimed Fabric Count:	/
Yarn Size:	/	Submitted Size:	/
Size Range:	/	FPU No.:	/
GPU No.:	/	End Use:	/
SKU:	/		

Product Category	1
Test Requested	PARTIAL TEST
Previous Report No.	

Submitted Fiber	70% PP + 30% PE
Content	
Submitted Care	1
Instruction(s)	
Client Expected Care	1
Instruction	
Suggested Care	1
Instruction(s)	

TEST PROPERTY	PASS	FAIL	DATA	COMMENTS
WATER RESISTANCE: IMPACT PENETRATION TEST	Х			
WATER RESISTANCE: HYDROSTATIC PRESSURE TEST	Х			

Revised Reason: As per vendor's request, delete some test in the report.

Remark: BVCPS (Shanghai) contact information for this report

Address: No.168, Guanghua Road, Zhuangiao Town, Minhang, Shanghai, China.

General Information:

Jojo Li 086-021-24081666*9309, email: jojo-wj.li@cn.bureauveritas.com Noah Zhang 086-021-24081666*1720, email: Xiaoliang.zhang@cn.bureauveritas.com Seven Sun 086-021-24081666*9327, email: Seven.sun@cn.bureauveritas.com Du Shuang 086-021-24081666*1709, email: shuang.du@cn.bureauveritas.com Challenge Qiao 086-021-24081666*9331, email: challenge.qiao@cn.bureauveritas.com Lizzy Liao 086-021-24081666*9303, email: lizzy.liao@bureauveritas.com

Sherry Shen

086-021-24081666*9310, email: shirley-ll.shen@cn.bureauveritas.com

Technical Questions:

Pat Liu 086-021-24081745, email: Pat.liu@cn.bureauveritas.com Rio Zhang 86-21-24081666*9305, email: Rio.Zhang@cn.bureauveritas.com

BUREAU VERITAS

CONSUMER PRODUCTS SERVICES (SHANGHAI)

Julia Yan

JULIA YAN

PROGRAM MANAGER

	TES	T RESULT	<u>'S</u>		
TEST PROPERTY					REQUIREMENTS
WATER RESISTANCE: IMPACT PENE	TRATION T	est (Aato	CC 42-17)		
	1	2	3	AVERAGE	
FABRIC (GRAM) SHOULDER (GRAM) ARM OPENING (GRAM) SLEEVE (GRAM)	0.0	0	0.0	0.0	MAX. 1.0 GRAM MAX. 1.0 GRAM MAX. 1.0 GRAM MAX. 1.0 GRAM
WATER RESISTANCE: HYDROSTATIC PRESSURE TEST (AATCC 127)					
	1	2	3	AVERAGE	
FABRIC (CMH ₂ O) SHOULDER (CMH ₂ O) ARM OPENING (CMH ₂ O) SLEEVE (CMH ₂ O)	170	3	171 2.8 9.6 5.2	173	MIN. 20 CMH ₂ O MIN. 20 CMH ₂ O MIN. 20 CMH ₂ O MIN. 20 CMH ₂ O



END OF REPORT